



INDIAN INSTITUTE OF TECHNOLOGY KANPUR
Department of Chemistry & Laser Technology Program

Dr. Debabrata Goswami, Professor

Post Office - I.I.T, Kanpur - 208016 (India)

August 14, 2012

Enquiry No : *EHM/DG/2012-13/10 DT 14/8/12*
 Last Date: 21st August, 2012 *extended upto 28th Aug 2012*

Subject: Request for Quotation for 'Motorized Linear Stages'

Dear Sir,

Please send sealed Quotation(s) with all technical details of:

Serial No.	Name	Quantity	Specification
1	Miniature linear stage	3	<ul style="list-style-type: none"> ➤ Travel Range 25 mm ➤ Resolution ~ 7.5 nm ➤ Pitch 200 or ± 100 μrad ➤ Load Capacity 50 N ➤ Normal Load Capacity (Cz) 50 N
2	Mid-Range Travel Linear Stage	2	<ul style="list-style-type: none"> ➤ Travel Range 150 mm ➤ Resolution 0.1 μm ➤ Pitch 140 or ± 70 μrad ➤ Load Capacity 200 N ➤ Normal Load Capacity (Cz) 200 N

Please send your technical and commercial offer on or before 21st August, 2012, to the following address:

Prof. D. Goswami
 Department of Chemistry
 IIT Kanpur
 Kanpur- 208016
 India

Offers can also be sent via email to: dgoswami@iitk.ac.in

Thanking you,

Regards,

Dr. D. Goswami
 Professor
 Dept. of Chemistry