

Enquiry No. IITK/CHM/PKB/2013-14/03

Date 14/05/2013

Kindly send us urgent sealed quotations for the following items. quotations should be addressed to the Prof. P. K. Bharadwaj, Chemistry Department, IIT Kanpur and must reach on or before Tuesday, the 21/05/2013.

Equipment	Description	Quantity
Rotary Evaporator	Specifications: 1. Motor Lift 2. Rotation Speed: 20-280 RPM 3. Rotation Speed setting: 4.3” LCD graphic display 4. Temperature Range: 20-210 °C 5. Temp accuracy: +/- 1°C 6. Bath Temp setting: 4.3” LCD graphic display 7. Heating bath Temperature control:electronic/digital 8. Volume of Heating bath: 4.5 lit. 9. Vapor temperature display 10. Integrated vacuum controller 11. Programmable ramps	1
Vacuum Vulve		
Vacuum Pump	Two-stage diaphragm pump made from chemically resistant materials 1. High suction capacity of 1.7 m ³ /hr. 2. Ultimate vacuum of 9 mbar 3. power input: 180 W 4. weight 10.5 kg Dimensions: 245X295X185 [L x W x H]	1

Prof. P. K. Bharadwaj
Head, Department of Chemistry
IIT Kanpur
Kanpur-208016

