

Department of Materials Science & Engineering

Indian Institute of Technology, Kanpur

Call for Quotation: Agilent Digital Multimeter 34401A

Enquiry No: MSE/SO/11-12/05

Enquiry Date: 4/1/2012

Closing Date: 13/1/2012

We are interested in the purchase of 'Agilent Digital Multimeter 34401A' for the four-point probe dc conductivity measurement. We are hereby calling for quotations from the prospective suppliers for the "Agilent Digital Multimeter" with the following minimum specifications/requirements:

1. Measure up to 1000 volts with 6.5 digits resolution
2. 0.0015% basic DC V accuracy (24 hour)
3. 0.06% basic AC V accuracy
4. 3 Hz to 300 kHz ac bandwidth
5. Can be connected to computer

Please send all technical specifications of the product being quoted and the quotation with the best price you can offer. You may highlight any special features of your product along with support evidence/data sheets.

Prospective suppliers may submit **Competitive quotation** for the 'Agilent DMM' in sealed envelope latest by **13th January, 2011** to the following address:

Dr. Shobit Omar
Materials Science & Engineering,
Indian Institute of Technology, Kanpur