



Enquiry No. : CHE/GD/BRNS/2011/01

December 9, 2011

Quotations are invited from interested suppliers for **Online UV Ozone Analyzer** for measurement of ozone concentration from 1-1000 ppm in a continuous flow system as per the following specifications.

Specifications

Ranges: Three (03) ranges: 0-1.0 ppm, 0-10 ppm, 0-900 ppm

Resolution: Should be less than 5% of the scale

Flow rate: Inlet flow rate should not exceed 1 LPM (less than 1 LPM will be preferred).

Pressure: Should take inlet gas at atmospheric conditions.

Response time: Should be less than about 1 minute.

Kindly mention about the terms and conditions such as;

- * Validity of the quotation for a minimum of 60 days.
- * Warranty period
- * Mention the FOB rate.
- * Site requirements; e.g. space, dimension and power requirement etc.
- * Special educational discounts, if any.

For indigenous or import items please use the form number 603 or 608 respectively given at

<http://web.iitk.ac.in/dord/rndforms/603.pdf>

<http://web.iitk.ac.in/dord/rndforms/608.pdf>

The quotations are to be sent to the under signed by December 24, 2011.

Thanking you;


Goutam Deo

Professor, Department of Chemical Engineering

CL 201B, IIT Kanpur

Kanpur- 208 016

Ph: 0512-2597881, Fax: 0512-2590104