

**Inquiry number: BSBE/SK/21-22/05**

Date: 10/03/2022

Closing date: 20/03/2022

**Description: Basic Power Supply for Gel Electrophoresis and transfer system**

**Address to:**

**Dr. Suresh Kumar**

Dept. of Biological Sciences and  
Bioengineering, Indian Institute of  
Technology, Kanpur  
Phone: 0512-2592247  
Email: [sureshkr@iitk.ac.in](mailto:sureshkr@iitk.ac.in)

**Product Specification:**

---

Output specifications

10–300 V, fully adjustable in 1 V steps

4–400 mA, fully adjustable in 1 mA steps

75 W (maximum)

Type of output: Constant voltage or constant current with automatic crossover

Output terminals: 4 pair recessed banana jacks in parallel

Timer control: 1 min–99 hr 59 min, fully adjustable

Pause/resume function: Yes

Display: 3-digit LED

Operating conditions: 0–40°C; 0–95% humidity in absence of condensation

Fuse on hot and neutral

Dimensions (W x D x H), cm/in

21 x 24.5 x 6.5/8.3 x 9.6 x 2.6